

Method for correcting calibration values in a calibration table of a computed tomography apparatus

Patent Number: ☐ US2002165686
Publication date: 2002-11-07
Inventor(s): KROPFELD HELMUT (DE); SCHWARZ KARL (US)
Applicant(s): SIEMENS AG (US)
Requested Patent: ☐ DE10112792
Application Number: US20020090655 20020305
Priority Number(s): DE20011012792 20010316
IPC Classification: G01D18/00; G01M19/00; G06F19/00; G01D21/00; G01P21/00; G01R35/00; G01C19/00; G01C25/00; G01F25/00
EC Classification: G01N23/04D
Equivalents: ☐ JP2002345806

Abstract

In a method for correcting a calibration table $T(n,k)$ of a CT apparatus that contains calibration values, the CT apparatus having a detector system with $N \geq 2$ rows of detector elements following one another in the z-direction that include a first active row of detector elements in the z-direction and a last active row of detector elements in the z-direction, for correcting the calibration values of the aforementioned first and last active rows of detector elements, a reference vector $R(k)$ is produced, the error is determined with respect to the first and last active rows of detector elements, and the error $F(n,k)$ of the first and last active rows of detector elements is subtracted from the corresponding calibration values of the calibration table $T(n,k)$ for determining corrected calibration values $T_{cor}(n,k)$ with respect to the first and last active rows of detector elements

Data supplied from the esp@cenet database - I2